

COOLEYSMILES
GENERAL & COSMETIC DENTISTRY
4100 Factoria Blvd. S.E., Suite C
Bellevue, WA 98006
425 747-7000
www.cooleysmiles.com

OCCLUSAL GUARD (NIGHT GUARD)

Every effort has been made to provide you with the most comfortable therapeutic dental appliance to meet your specific treatment objectives. Proper care and maintenance by you at home will help insure that the appliance will continue to function properly and minimize any deterioration of its excellent qualities.

1. Never allow the appliance near high temperatures nor allow prolonged dehydration.
2. When the appliance is in regular use, allow it to air dry after cleaning until next use. If the appliance will not be worn for periods of time exceeding one week, it should be stored in a humidior. A seal able Tupperware type container with a wet paper towel is suitable.
3. Soften the appliance under warm tap water prior to placement.
4. Harden the appliance under cold tap water prior to cleaning using a soft brush if necessary to remove debris.
5. If mineral deposits or discoloration appear, or you have difficulty keeping the appliance clean, contact your dentist whose laboratory can provide the appropriate cleanser.
6. In normal use, the appearance of the retentive portion may vary from crystal clear to mildly opaque.
7. If you notice any changes in your symptoms, appearance of your mouth or gums, or changes in the appliance itself, contact your dentist.
8. Removal of the appliance is best accomplished by using equal pull on both sides of your mouth. This will minimize the chance of damage to the resilient portion of the appliance.

Thank you,

Dr. Aaron C. Cooley and Dr. Brandon E. Cooley,